

Sr. No.	Date	Dept	Description of Complaint	Wdy. / Date Taken	Sign of Technical person / Incharge	Sign of Department Representative	Sign of Maintenance Committee / Review
1	29/11/16	Endo	right cure unit is not working	21/11/17 repaired	<u>Dr. Rajkumar</u> 18/11/17	Dr. Rajkumar	Dr. Savita 30/11/17
2	01/12/16	Radio	2 ultrasonic scales do not functional	21/11/17 repaired	<u>Dr. Patel</u> 18/11/17		
3	05/12/16	Endo	lecture hall tube light is not working	21/11/17 repaired	<u>Dr. Patel</u> 18/11/17		
4	30/11/17	O.S	front loading machine is not working	20/12/17 repaired	<u>Dr. Patel</u> 24/11/17	Dr. Trupti P	Dr. Savita 27/11/17
5	15/11/17	Endo	2 dental chairs are not functional	20/11/17 repaired	<u>Dr. Patel</u> 24/11/17	Dr. Beelma	
6	30/11/17		condenser unit signal provided		<u>Dr. Patel</u> 24/11/17	NA present	

M. J. Vandeekar

DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

Sr. No	Date	Dept	Description of Complaint	Action taken	Signature of Technical Person in charge	Signature of Department Representative	Signature of Matricular Committee (Review)
1	10/4/17	ONOR	X-ray machine not functioning	Technician replaced 25/11/17	<i>P. J. J.</i>	Dr. Subhashini	Dr. Trupti 31/7/17
2	17/5/17	Medico	2 Anesthesia & not working	1 Cassing replaced 25/11/17 1 Dr. J. J. replaced 25/11/17	<i>P. J. J.</i>	Dr. Narasimha	Dr. Trupti 31/7/17
3	23/6/17	Library	Binding of Books	Binding of Book done 25/11/17	<i>U. J. J.</i>	Mr. V. S. S.	
4	19/7/17	PHD	Light Cure unit is not functioning	Replacement of bulb in the unit is done	<i>P. J. J.</i>	Dr. B. R. V.	
5	18/8/17	Pdco	Compressor of 3 dental chair & not functioning	Oil change has been done & compressor is working properly	<i>P. J. J.</i>	Dr. P. S. J.	Dr. Trupti 28/10/17
6	25/9/17	Avulsi	AV System is not working	Repair done	<i>P. J. J.</i>	Mr. P. S. J.	

M. J. J.

DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

	Sl. No.	Date	Description of Complaint	Ref. Name	Ref. Address	Ref. Phone
1	1	21/11/17	Implant physidymia is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
2	2	21/11/17	Rug is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
3	3	21/11/17	Generator is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
4	4	21/11/17	Miscellaneous is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
5	5	21/11/17	Opq Machine is not functioning	Dr. J. J. J.	Dr. J. J. J.	24/11/17
6	6	21/11/17	Mostly weeds burning	Dr. J. J. J.	Dr. J. J. J.	24/11/17
7	7	21/11/17	Generator is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
8	8	21/11/17	Implant physidymia is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
9	9	21/11/17	Rug is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
10	10	21/11/17	Generator is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
11	11	21/11/17	Miscellaneous is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
12	12	21/11/17	Opq Machine is not functioning	Dr. J. J. J.	Dr. J. J. J.	24/11/17
13	13	21/11/17	Mostly weeds burning	Dr. J. J. J.	Dr. J. J. J.	24/11/17
14	14	21/11/17	Generator is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
15	15	21/11/17	Implant physidymia is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
16	16	21/11/17	Rug is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
17	17	21/11/17	Generator is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
18	18	21/11/17	Miscellaneous is not working	Dr. J. J. J.	Dr. J. J. J.	24/11/17
19	19	21/11/17	Opq Machine is not functioning	Dr. J. J. J.	Dr. J. J. J.	24/11/17
20	20	21/11/17	Mostly weeds burning	Dr. J. J. J.	Dr. J. J. J.	24/11/17

M. J. Vandeekar

DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

12	18/7/18	PHD	On study is not functional	Water Chutney is served	<i>[Signature]</i>	Dr. Renu	
13	11/6/18	Padi	Leakage in Sanitary	Water spraying	<i>[Signature]</i>	Ms. Shweta	
14	5/5/18	Other	Upper Cus Unit is not functional	For change in the unit	<i>[Signature]</i>	Dr. Disha	
15	11/6/18	Padi	Problems in not working	Problems in working properly	<i>[Signature]</i>	Dr. Sunil	
16	12/9/18	6 Hospital	In General ward 3 fans are not working	Capacitor is replaced	<i>[Signature]</i>	Dr. Prachi	
17	23/8/18	OSD	Centrifuge machine is not functional	Machine is repaired	<i>[Signature]</i>	Dr. Prachi	
18	12/9/18	6 Hospital	In General ward 3 fans are not working	Capacitor is replaced	<i>[Signature]</i>	Dr. Prachi	

M. J. Vandeekar


DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210






Ser	Date	Dpt	Description of Complaint	Actual	Exam	Sign of infection	Sign of dysentery	Sign of malocclusion
25	22/11/18	dentistry	Dissection had ligatures were missing	Teeth were replaced	Post	Dr Shukla	Dr. C. K. K.	22/01/19
26	16/11/18	O.S	Boyle's separation is not in position	Lights have been replaced	Post	Dr. P. P. P.	Dr. P. P. P.	
27	9/12/18	lecture hall	Projector is not working	Replacement of focus lens	Post	Mr. P. P. P.		
28	17/11/19	G medicine	B.P apparatus is not working	Cuff has been replaced	Post	Dr. S. S. S.		
29	29/12/19	Pedio	Front hooding is not working	Trimmed to be adjusted	Post	Dr. S. S. S.		23/04/19
30	13/3/19		Light 8 one from 20 not working in lobby area	Lights are from have been repaired	Post	Mr. P. P. P.		


M. J. Kulkarni

DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

Sino Date	Rpt	Description of Complaint	How taken	dent / lesion	Tech / sign of damage	sign of department	sign of maintenance
31 11/11/19	C Patient	2 microscope lens is damaged	both lenses have been replaced		Dr. Devesh	Dr. Devesh	Dr. Savita 24/09/19
30 15/11/19	G Surgery	O.T lights are not working	lights & bulbs were replaced		Dr. Pramod		
29 30/11/19	Garden	Pipe water in garden	leakage problem solved		Ms. Raju		
24 16/11/19	Physiology	Julia Com is not working	Cable has been replaced		Mr. Nilesh		
35 19/11/19	Endo	2 Hualon Hand piece is not working	Changing, carbide of end, polishing of end has been done		Dr. Rajkumar		Dr. Savita 26/10/19
36 18/11/19	Reception	Dut laptop is not working	CPU fan has been repaired		Mr. Zaveri		


DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

Ser No	Date	Dept	Description of Complaint	Action	done	Signature of Technical person in Charge	Signature of Department Representative	Signature of Managing Committee President
37	15/10/19	omde	3 dental Chair again are not working	3 dental Chair have been fixed			Dr. Sanita	Dr. Bindus 90/10/20
38	25/11/19	Ors	Attache case is burnt	case has been replaced			Dr. Tusha	
39	20/12/19	Spots	T.T Table broken	T.T table has been repaired			Mr. Shyam Kulkarni	
40	20/11/20	Paid	Physiotherapy is not working	Physiotherapy has been repaired			Dr. Sumana	
41	19/12/20	Endo	Dr's machine is not working	Machine, being changed			Dr. Raji	Dr. Bindus 23/01/20
42	16/3/20		Dental Chair CCN is not working	Cable of chair has been changed				


DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210

Sl. no	Date	Dept	Description of Complaint	Use of Machine / Action	Sign of Technical Person in Charge	Sign of Representative	Sign of Maintaining Committee
49	22/10/20	Prosthodontics	Upper lower denture to be repaired	Tip of the denture replaced	<u>P. N. D. J.</u>	Dr. Rajana	Dr. Divya 27/11/21
50	21/11/20	Endodontics	Teeth machine is not working	Teeth machine is repaired	<u>P. N. D. J.</u>	Dr. Nalini	Dr. Divya 27/11/21
51	21/11/20	Fixed	Fixed appliance is not working	Fixed appliance is repaired	<u>P. N. D. J.</u>	Mr. Sukar	
52	18/11/21	Pedodontics	Loose machine is not working	Loose machine is repaired	<u>P. N. D. J.</u>	<u>P. N. D. J.</u>	
53	15/11/21	Orthodontics	Upper lower denture is not working	Upper lower denture is repaired	<u>P. N. D. J.</u>	Dr. Ruma	Dr. Divya 26/11/21
54	22/11/21	Orthodontics	Denture is not functional	Denture is repaired	<u>P. N. D. J.</u>	Dr. Ruma	

M. J. Vandeekar

DEAN
Y.M.T. Dental College
& Hospital Kharghar,
Navi Mumbai - 410 210